- RAISON, J.C.A., BEAUMONT, J.O., RUSSELL, J.A.G., OSBORN, J.J. & GERBODE, F. (1968) Alarms in an intensive care unit: an interim compromise. *Computers and Biomedical Research*, **1**, 556.
- REPORT (1969) Brussels exhibition 1969. *The Inventor*, **9**, 26.
- SADERA, T.D., GURUSWAMY, B.A., PENNY, R.K., STEWART J.S.S. (1969) *Design of a circulatory failure analyser*. (In press.)
- STEWART, J.S.S. (1965) Advances in the management of cardiac arrest. *Journal of the Royal College of Surgeons (Edinburgh)*, **10**, 85.
- STEWART, J.S.S. (1969a) Meaningful monitoring. *Lancet*, **i**, 1305.
- STEWART, J.S.S. (1969b) A clinical diagnostic computer and monitor. *European Surgical Research*, **1**, 248.
- STEWART, J.S.S. & BARDHAN, K.D. (1966) Intensive care units for myocardial infarction. *British Medical Journal*, **i**, 1297.
- TAYLOR, D.J.W. (1968) The medical engineering interface in Britain. *Instrumentation in Medicine*. Proceedings of Second European Symposium on medical electronics, London 1967. Hanover Press, London.
- WALLER, A.D. (1887) A demonstration on man of electromotive changes accompanying the heart's beat. *Journal of Physiology*, **8**, 229.
- WOLFF, H.S. (1968) Bioengineering—the next few years. *Bio-Medical Engineering*, **3**, 15.

Sensors

H. S. WOLFF

Medical Research Council, Hampstead Laboratories, Holly Hill, London

(Abstract)

ANY CHANNEL of a monitoring system, however simple, must consist at least of a sensor or transducer, a signal processor and a display device. Of these three components, signal processing which is usually a purely electronic function presents the smallest problems, display devices are adequate at the present state of the art, but the acquisition of the input signal can still present considerable difficulties.

These difficulties can be classified under three headings. First, the choice of the signal to be sensed. For reasons of patient comfort, convenience and safety, signals are often chosen which may only be remotely connected to the physiological process which is to be monitored. An example is the indirect

measurement of arterial blood pressure, which however good the sensing system may be, will always be inherently more unreliable than direct measurement. Second, the stability of the transducer itself, problems of calibration and sterilization and resistance to maltreatment and overload. Thirdly, methods of attachment or connection to the patient such that voluntary and involuntary movements cause the minimum of artifact and present the least obstacle to nursing procedure.

Sensors in present day use and some which are under development were discussed with reference to the above classifications.